

United States Patent and Trademark Office

UNITED STATES DEPARTMENT OF COMMERCE United States Patent and Trademark Office Address: COMMISSIONER FOR PATENTS P.O. Box 1450 Alexandria, Virginia 22313-1450 www.uspto.gov

APPLICATION NO.	FILING DATE	FIRST NAMED INVENTOR	ATTORNEY DOCKET NO.	CONFIRMATION NO.
10/536,798	05/27/2005	Toru Sano	Q88016	4489
23373 7590 10/29/2007 SUGHRUE MION, PLLC			EXAMINER	
2100 PENNSY	LVANIA AVENUE, N	.W.	THERKORN, ERNEST G	
SUITE 800 WASHINGTON, DC 20037			ART UNIT	PAPER NUMBER
			1797	
•				
			MAIL DATE ·	DELIVERY MODE
	•		10/29/2007	PAPER

Please find below and/or attached an Office communication concerning this application or proceeding.

The time period for reply, if any, is set in the attached communication.

•						
	Application No.	Applicant(s)				
	10/536,798	SANO ET AL.				
Office Action Summary	Examiner	Art Unit				
	Ernest G. Therkorn	1797				
The MAILING DATE of this communication Period for Reply	n appears on the cover sheet wit	th the correspondence address				
A SHORTENED STATUTORY PERIOD FOR R WHICHEVER IS LONGER, FROM THE MAILIN - Extensions of time may be available under the provisions of 37 C after SIX (6) MONTHS from the mailing date of this communication. If NO period for reply is specified above, the maximum statutory provides the second of th	IG DATE OF THIS COMMUNIC FR 1.136(a). In no event, however, may a re on. period will apply and will expire SIX (6) MONT statute, cause the application to become ABA	CATION. Apply be timely filed THS from the mailing date of this communication. ANDONED (35 U.S.C. § 133).				
Status						
1) Responsive to communication(s) filed on	12 October 2007.					
2a) This action is FINAL . 2b)⊠	This action is FINAL . 2b)⊠ This action is non-final.					
3) Since this application is in condition for allowance except for formal matters, prosecution as to the merits is						
closed in accordance with the practice un	der <i>Ex parte Quayle</i> , 1935 C.D.	. 11, 453 O.G. 213.				
Disposition of Claims						
4) ☐ Claim(s) 1-9 is/are pending in the applicat 4a) Of the above claim(s) 7-9 is/are withdr 5) ☐ Claim(s) is/are allowed. 6) ☐ Claim(s) 1-6 is/are rejected. 7) ☐ Claim(s) is/are objected to. 8) ☐ Claim(s) are subject to restriction a	awn from consideration.					
Application Papers						
9) ☐ The specification is objected to by the Exa	miner.					
10) The drawing(s) filed on is/are: a) □ accepted or b) □ objected to by the Examiner.						
Applicant may not request that any objection to	o the drawing(s) be held in abeyan	ce. See 37 CFR 1.85(a).				
Replacement drawing sheet(s) including the control of the control		· · · · · · · · · · · · · · · · · · ·				
Priority under 35 U.S.C. § 119						
12) Acknowledgment is made of a claim for for a) All b) Some * c) None of: 1. Certified copies of the priority docur 2. Certified copies of the priority docur 3. Copies of the certified copies of the application from the International But * See the attached detailed Office action for a	ments have been received. ments have been received in Appriority documents have been ureau (PCT Rule 17.2(a)).	pplication No received in this National Stage				
Attachment(s)	_					
 Notice of References Cited (PTO-892) Notice of Draftsperson's Patent Drawing Review (PTO-94 Information Disclosure Statement(s) (PTO/SB/08) Paper No(s)/Mail Date 	8) Paper No(s	ummary (PTO-413))/Mail Date Iformal Patent Application 				

Art Unit: 1797

The following is a quotation of the appropriate paragraphs of 35 U.S.C. 102 that form the basis for the rejections under this section made in this Office action:

A person shall be entitled to a patent unless -

- (b) the invention was patented or described in a printed publication in this or a foreign country or in public use or on sale in this country, more than one year prior to the date of application for patent in the United States.
- (e) the invention was described in (1) an application for patent, published under section 122(b), by another filed in the United States before the invention by the applicant for patent or (2) a patent granted on an application for patent by another filed in the United States before the invention by the applicant for patent, except that an international application filed under the treaty defined in section 351(a) shall have the effects for purposes of this subsection of an application filed in the United States only if the international application designated the United States and was published under Article 21(2) of such treaty in the English language.

The following is a quotation of 35 U.S.C. 103(a) which forms the basis for all obviousness rejections set forth in this Office action:

(a) A patent may not be obtained though the invention is not identically disclosed or described as set forth in section 102 of this title, if the differences between the subject matter sought to be patented and the prior art are such that the subject matter as a whole would have been obvious at the time the invention was made to a person having ordinary skill in the art to which said subject matter pertains. Patentability shall not be negatived by the manner in which the invention was made.

Claims 1 and 2 are rejected under 35 U.S.C. 102(B and/or E) as anticipated by or, in the alternative, under 35 U.S.C. 103(a) as obvious over each of Christel (WO/99/09042), Hansmann (U.S. Patent No. 5,952,173), and Moon (U.S. Patent Publication No. 2002/0123153). The claims are considered to read on each of Christel (WO/99/09042), Hansmann (U.S. Patent No. 5,952,173), and Moon (U.S. Patent Publication No. 2002/0123153). However, if a difference exists between the claims and each of Christel (WO/99/09042), Hansmann (U.S. Patent No. 5,952,173), and Moon (U.S. Patent Publication No. 2002/0123153), it would reside in optimizing the elements of each of Christel (WO/99/09042), Hansmann (U.S. Patent No. 5,952,173), and Moon (U.S. Patent Publication No. 2002/0123153). It would have been obvious to optimize

Art Unit: 1797

the elements of each of Christel (WO/99/09042), Hansmann (U.S. Patent No. 5,952,173), and Moon (U.S. Patent Publication No. 2002/0123153) to enhance separation.

Claims 3 and 4 are rejected under 35 U.S.C. 103(a) as being unpatentable over each of Christel (WO/99/09042), Hansmann (U.S. Patent No. 5,952,173), and Moon (U.S. Patent Publication No. 2002/0123153) in view of Nagaoka (WO/02/23180). At best, the claims differ from each of Christel (WO/99/09042), Hansmann (U.S. Patent No. 5,952,173), and Moon (U.S. Patent Publication No. 2002/0123153) in reciting use of electrodes. Nagaoka (WO/02/23180) (Abstract) discloses that use of electrodes allows coupling of the specific component to a projection. It would have been obvious to use electrodes in each of Christel (WO/99/09042), Hansmann (U.S. Patent No. 5,952,173), and Moon (U.S. Patent Publication No. 2002/0123153) because Nagaoka (WO/02/23180) (Abstract) discloses that use of electrodes allows coupling of the specific component to a projection.

Claim 5 is rejected under 35 U.S.C. 103(a) as being unpatentable over each of Christel (WO/99/09042), Hansmann (U.S. Patent No. 5,952,173), and Moon (U.S. Patent Publication No. 2002/0123153) in view of either Anderson (U.S. Patent No. 6,289,286) or Malmqvist (U.S. Patent No. 6,503,760). At best, the claim differs from each of Christel (WO/99/09042), Hansmann (U.S. Patent No. 5,952,173), and Moon (U.S. Patent Publication No. 2002/0123153) in reciting use of an antigen-antibody combination. Anderson (U.S. Patent No. 6,289,286) (column 1, lines 18-34 and column 23, line 7) and Malmqvist (U.S. Patent No. 6,503,760) (column 1, lines 19-34. column 6,

Art Unit: 1797

line 60) disclose antibody-antigen interaction with a surface bound ligand are of fundamental importance in many fields and are used in microfluidic analytical instruments. It would have been obvious to use an antigen-antibody combination in each of Christel (WO/99/09042), Hansmann (U.S. Patent No. 5,952,173), and Moon (U.S. Patent Publication No. 2002/0123153) in view of either Anderson (U.S. Patent No. 6,289,286) or Malmqvist (U.S. Patent No. 6,503,760) because Anderson (U.S. Patent No. 6,289,286) (column 1, lines 18-34 and column 23, line 7) and Malmqvist (U.S. Patent No. 6,503,760) (column 1, lines 19-34. column 6, line 60) disclose antibody-antigen interaction with a surface bound ligand are of fundamental importance in many fields and are used in microfluidic analytical instruments.

Claim 6 is rejected under 35 U.S.C. 103(a) as being unpatentable over each of Christel (WO/99/09042), Hansmann (U.S. Patent No. 5,952,173), and Moon (U.S. Patent Publication No. 2002/0123153) in view of Biebricher (U.S. Patent No. 4,177,038). At best, the claim differs from each of Christel (WO/99/09042), Hansmann (U.S. Patent No. 5,952,173), and Moon (U.S. Patent Publication No. 2002/0123153) in reciting use of a spacer. Biebricher (U.S. Patent No. 4,177,038) (column 1, lines 34-40) discloses use of a spacer increases the distance between the vehicle and the bonding material to counteract steric interference. It would have been obvious to use a spacer in each of Christel (WO/99/09042), Hansmann (U.S. Patent No. 5,952,173), and Moon (U.S. Patent Publication No. 2002/0123153) because Biebricher (U.S. Patent No. 4,177,038) (column 1, lines 34-40) discloses use of a spacer increases the distance between the vehicle and the bonding material to counteract steric interference.

Art Unit: 1797

Claims 1-4 are rejected under 35 U.S.C. 102(B and/or E) as anticipated by or, in the alternative, under 35 U.S.C. 103(a) as obvious over Nagaoka (WO/02/23180). The claims are considered to read on Nagaoka (WO/02/23180). However, if a difference exists between the claims and Nagaoka (WO/02/23180), it would reside in optimizing the elements of Nagaoka (WO/02/23180). It would have been obvious to optimize the elements of Nagaoka (WO/02/23180) to enhance separation.

Claim 5 is rejected under 35 U.S.C. 103(a) as being unpatentable over Nagaoka (WO/02/23180) in view of either Anderson (U.S. Patent No. 6,289,286) or Malmqvist (U.S. Patent No. 6,503,760). At best, the claim differs from Nagaoka (WO/02/23180) in reciting use of an antigen-antibody combination. Anderson (U.S. Patent No. 6,289,286) (column 1, lines 18-34 and column 23, line 7) and Malmqvist (U.S. Patent No. 6,503,760) (column 1, lines 19-34. column 6, line 60) disclose antibody-antigen interaction with a surface bound ligand are of fundamental importance in many fields and are used in microfluidic analytical instruments. It would have been obvious to use an antigen-antibody combination in Nagaoka (WO/02/23180) because Anderson (U.S. Patent No. 6,289,286) (column 1, lines 18-34 and column 23, line 7) and Malmqvist (U.S. Patent No. 6,503,760) (column 1, lines 19-34. column 6, line 60) disclose antibody-antigen interaction with a surface bound ligand are of fundamental importance in many fields and are used in microfluidic analytical instruments.

Claim 6 is rejected under 35 U.S.C. 103(a) as being unpatentable over Nagaoka (WO/02/23180) in view of either Anderson (U.S. Patent No. 6,289,286) or Malmqvist (U.S. Patent No. 6,503,760) as applied to claim 5 above, and further in view of

Art Unit: 1797

Biebricher (U.S. Patent No. 4,177,038). At best, the claim differs from Nagaoka (WO/02/23180) in view of either Anderson (U.S. Patent No. 6,289,286) or Malmqvist (U.S. Patent No. 6,503,760) in reciting use of a spacer. Biebricher (U.S. Patent No. 4,177,038) (column 1, lines 34-40) discloses use of a spacer increases the distance between the vehicle and the bonding material to counteract steric interference. It would have been obvious to use a spacer in Nagaoka (WO/02/23180) in view of either Anderson (U.S. Patent No. 6,289,286) or Malmqvist (U.S. Patent No. 6,503,760) because Biebricher (U.S. Patent No. 4,177,038) (column 1, lines 34-40) discloses use of a spacer increases the distance between the vehicle and the bonding material to counteract steric interference.

The restriction requirement has been reconsidered, deemed proper, and made final for the reasons of record.

Any inquiry concerning this communication should be directed to E. Therkorn at telephone number (571) 272-1149. The official fax number is 571-273-8300.

Information regarding the status of an application may be obtained from the Patent Application Information Retrieval (PAIR) system. Status information for published applications may be obtained from either Private PAIR or Public PAIR. Status information for unpublished applications is available through Private PAIR only. For more information about the PAIR system, see http://pair-direct.uspto.gov. Should you have guestions on access to the Private PAIR system, contact the Electronic

Art Unit: 1797

Business Center (EBC) at 866-217-9197 (toll-free).

Ernest G. Therkorn Primary Examiner Art Unit 1797

Errot 6 thelon

Page 7

EGT October 25, 2007